Radioactive Material Inventory and Leak Test Form

36th Civil Engineering Squadron (EM)

Permit #: Generally Licensed Material

Date: 15-Dec-10

Inventory Frequency: ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

Leak Test Frequency: ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

Check One: ☑ Inventory ☑ Inventory and Leak Test

Source Information						Leak Test Information[1]		
Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location	Swipe Sample #	Date Collected	Result
Ni-63	10 mCi	APD 2000	5691		CEX Warehouse	WW120001	15-Dec-10	
Ni-63	10 mCi	APD 2000	5690		CEX Warehouse	WW120001	15-Dec-10	
Ni-63	10 mCi	APD 2000	5692		CEX Warehouse	WW120001	15-Dec-10	
Ni-63	10 mCi	APD 2000	3128		CEX Warehouse	WW120001	15-Dec-10	

[1]Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

☑ All items accounted for

☐ All swipe results less than or equal to 0.005 uCi

Survey Accomplished By:	Reviewed By:
Name: MARCUS A. POULOS, SSgt, USAF, 4B071	MICHAEL P. MORAN, Maj, USAF, BSC
Signature:	