

## Radioactive Material Inventory and Leak Test Form

### Postal Service Center

Permit #: **Generally Licensed Material**

Inventory Frequency:  Quarterly  Semi-Annual  Annual  Other \_\_\_\_\_

Leak Test Frequency:  Quarterly  Semi-Annual  Annual  Other \_\_\_\_\_

**Source Information**

Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location
Ni-63	10 mCi	IonScan 400	13143		Bldg 21001, Post Office

**Leak Test Information<sup>1</sup>**

Swipe Sample #	Date Collected	Result
WW120063	11-Dec-12	

<sup>1</sup>Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

- All items accounted for
- All swipe results less than or equal to 0.005 uCi

Date: 11 Dec 12  
Date Results Received:

**Survey Accomplished By:**

Name: **KRISTIAN M. ATKINSON, SSgt, USAF, 4B071**

Signature:

**Reviewed By:**

**KHAI H. VUONG, Maj, USAF, BSC**