Radioactive Material Inventory and Leak Test Form

Emergency Management

Permit #: Generally Licensed Material

Inventory Frequency: ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

Leak Test Frequency: ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

Source Information						Leak Test Information[1]		
Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location	Swipe Sample #	Date Collected	Result
Ni-63	10 mCi	APD 2000	5691		CEX Warehouse	WW-131202	17-Dec-13	
Ni-63	10 mCi	APD 2000	5690		CEX Warehouse	WW-131203	17-Dec-13	
Ni-63	10 mCi	APD 2000	5692		CEX Warehouse	WW-131204	17-Dec-13	
Ni-63	10 mCi	APD 2000	3123		CEX Warehouse	WW-131201	17-Dec-13	

[1]Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

☑ All items accounted for

☐ All swipe results less than or equal to 0.005 uCi

Date: 23 Dec 13

Date Results Received:

Survey Accomplished By:	Reviewed By:
Name: JEREMY A. BOLDT, SSgt, USAF, 4B071	GREGORY J. LATHBURY, Capt, USAF, BSC
Signature:	