# Radioactive Material Inventory and Leak Test Form

## Fire Department

**Permit #:** **Generally Licensed Material**

**Inventory Frequency:** ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

**Leak Test Frequency:** ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

| Source Information | | | | | | Leak Test Information[1] | | |
|---|---|---|---|---|---|---|---|---|
| **Isotope** | **Activity** | **Item Description** | **Unit Serial #** | **Source Serial #** | **Storage Location** | **Swipe Sample #** | **Date Collected** | **Result** |
| **Ni-63** | **10 mCi** | **APD 2000** | **2965** | | **Bldg 17002 Fire Station** | **WW-131208** | **17-Dec-13** | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |

[1]Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

☑ All items accounted for

☐ All swipe results less than or equal to 0.005 uCi

Date: 23 Dec 13

Date Results Received:

| **Survey Accomplished By:** | **Reviewed By:** |
|---|---|
| **Name:** **JEREMY A. BOLDT, SSgt, USAF, 4B071** | **GREGORY J. LATHBURY, Capt, USAF, BSC** |
| **Signature:** | |