

Radioactive Material Inventory and Leak Test Form

Fire Department

Permit #: **Generally Licensed Material**

Inventory Frequency: Quarterly Semi-Annual Annual Other _____

Leak Test Frequency: Quarterly Semi-Annual Annual Other _____

Source Information

Leak Test Information¹

Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location	Swipe Sample #	Date Collected	Result
Ni-63	10 mCi	APD 2000	2965		Bldg 17002 Fire Station	WW-141201	16-Dec-14	

¹Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

- All items accounted for Date: 16 Dec 14
- All swipe results less than or equal to 0.005 uCi Date Results Received:

Survey Accomplished By:

Reviewed By:

Name: **JEREMY A. BOLDT, SSgt, USAF, 4B071**

GREGORY J. LATHBURY, Capt, USAF, BSC

Signature: 

