Emergency Management Permit #: Generally Licensed Material			
Permit #: Generally Licensed Material			
Inventory Frequency:			
Leak Test Frequency:     ☐ Quarterly     ☑ Semi-Annual     ☐ Annual     ☐ Other			
Source Information  Leak Test Informatio  Isotope Activity Item Description Unit Serial # Source Serial # Storage Location Swipe Sample #	on <sup>1</sup> Date Collected	Result	
Ni-63 10 mCi APD 2000 3123 CEX Warehouse WW-140605	11-Jun-14		
Ni-63 10 mCi APD 2000 5691 CEX Warehouse WW-140606	11-Jun-14		
Ni-63 10 mCi APD 2000 5690 CEX Warehouse WW-140607	11-Jun-14		
Ni-63 10 mCi APD 2000 5692 CEX Warehouse WW-140608	11-Jun-14		
<sup>1</sup> Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.			
✓ All items accounted for Date: 13 Jun 14  ☐ All swipe results less than or equal to 0.005 uCi  Date Possilts Possilted:			
Date Nesults Neceived.			
Survey Accomplished By:  Name:  Reviewed By:			
JEREMY A. BOLDT, SSgt, USAF, 4B071 GREGORY J. LATHBURY, Capt, USAF, BSC Signature:	GREGORY J. LATHBURY, Capt, USAF, BSC		