Radioactive Material Inventory and Leak Test Form

Fire Department

Permit #: **Generally Licensed Material**

Inventory Frequency: ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

Leak Test Frequency: ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

Source Information						Leak Test Information[1]		
Isotope	**Activity**	**Item Description**	**Unit Serial #**	**Source Serial #**	**Storage Location**	**Swipe Sample #**	**Date Collected**	**Result**
Ni-63	**10 mCi**	**APD 2000**	**2965**		**Bldg 17002 Fire Station**	**WW-150601**	**30-Jun-15**	

[1]Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

☑ All items accounted for

☐ All swipe results less than or equal to 0.005 uCi

Date: 30 Jun 15

Date Results Received:

Survey Accomplished By:	**Reviewed By:**
Name: **JEREMY A. BOLDT, SSgt, USAF, 4B071**	**GREGORY J. LATHBURY, Capt, USAF, BSC**
Signature:	