

Radioactive Material Inventory and Leak Test Form

Postal Service Center

Permit #: **Generally Licensed Material**

Inventory Frequency: Quarterly Semi-Annual Annual Other _____

Leak Test Frequency: Quarterly Semi-Annual Annual Other _____

Source Information

Leak Test Information¹

Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location	Swipe Sample #	Date Collected	Result
Ni-63	10 mCi	IonScan 400	13143		Bldg 21001, Post Office	N/A		

¹Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

- All items accounted for
- All swipe results less than or equal to 0.005 uCi

Date: 16 Dec 14
Date Results Received:

Survey Accomplished By:

Reviewed By:

Name: JEREMY A. BOLDT, SSgt, USAF, 4B071

GREGORY J. LATHBURY, Capt, USAF, BSC

Signature: 

